



Blood Glucose  
Measurement

# Omnitest<sup>®</sup> 3

COULD IT BE EASIER?

ISO  
15197:2013

# Omnitest<sup>®</sup> 3

COULD IT BE EASIER?



## NO CODING

After strip insertion, Omnitest® 3 is ready for the measurement. The code is recognized automatically. The intuitive menu navigation ensures a simple and safe handling.



## ONLY 0.3 µL SAMPLE VOLUME

Omnitest® 3 gets along with a minimum blood quantity of 0.3 µL. The test strip absorbs the blood automatically and starts the measurement as soon as a sufficient sample quantity is available.



## 3, 2, 1 – READY

In only 3 seconds Omnitest® 3 provides reliable results. No other blood glucose monitoring system is faster. The measurement result is displayed – depending on the type of device – in mg/dL or mmol/L.



# Omnitest<sup>®</sup> 3

COULD IT BE EASIER?



Image: original size

## SAFETY WITH A SINGLE CLICK

The convenient strip ejector allows for a hygienical test strip removal without touching it. A helpful feature – particularly if family members or professionals perform the measurement.



## MARKING FUNCTION

To evaluate the effect of nutrition and medication on the blood glucose level, Omnitest® 3 provides a marking function. If required, a measurement result could be stored together with the chosen marker.



## DATA ANALYSIS

Omnitest® 3 stores up to 365 results incl. date and time and calculates three different average values. All stored data can be transferred to a PC via an interface cable.



# Omnitest<sup>®</sup> 3

COULD IT BE EASIER?



---

Omnitest<sup>®</sup> 3 is available in a convenient pouch – so that nothing gets lost on the way

---



- Auto on with strip insertion
- Auto off in 3 minutes when meter is not used
- Audible signal when sufficient sample is applied
- Display of the temperature
- 5 customizable alarms
- Calculation of 3 average values in the range 1 to 99 days can be selected
- Possibility to delete values
- Marking function for pre- and postmeal, after taking medication and control tests
- Pouch included, meter could be used without being removed from the pouch
- Rubber feet for stable stand

#### PLEASE NOTE

With Omnitest® 3, the measurement unit mg/dL or mmol/L is prefixed!

# Omnitest<sup>®</sup> 3

## AT A GLANCE

Sample volume	0.3 µL
Measuring time	3 seconds
Code	Automatic code recognition
Measuring range	0.6 – 33.3 mmol/L
Operating temperature	10 °C – 40 °C or 50 °F – 104 °F
Operating humidity	10 – 90%
Hematocrit	20 – 60% (0.6 – 8.3 mmol/L) 30 – 55% (8.3 – 33.3 mmol/L)
Calibration	Plasma
Memory capacity	365 test results incl. date and time
Enzyme	Glucose oxidase
Dimensions	84 x 56 x 18 mm (L x W x H)
Weight	54 g (including batteries)
Power supply	2 x 3V Lithium batteries (CR2032)
Battery life	Minimum 3,000 tests
Data transfer to a computer	DIABASS <sup>®</sup> software (mediaspects GmbH), PC-cable is required. More info: <a href="http://www.bbraun.com/diabetes-datamanagement">www.bbraun.com/diabetes-datamanagement</a>
Fulfills the requirements of	ISO 15197:2013 and EN ISO 15197:2015

B. Braun Melsungen AG | [www.bbraun.com/diabetes](http://www.bbraun.com/diabetes)