

Omnitest® plus

CHECKLIST: ACCURACY COMPLAINTS

PROBLEM:

ASSUMED ABNORMAL VALUE
(CHAPTER A, B, C, D)

DIFFERENT VALUE THAN REFERENCE DEVICE
(CHAPTER E, F)


HAVE READY:

Device, a test strip from the same vial as the used test strip, check strip and control solution Omnitest® plus Control.

Omnitest® plus
serial number **SN**




Omnitest® plus
test strip lot **LOT**



PROBLEM: ASSUMED ABNORMAL VALUE

A: HANDLING CHECK



Did you set the code as indicated
on your test strip vial?

NO

YES

Did you already perform
a measurement?

YES

NO



Please re-insert the
test strip and set
the code as indicated
on your test strip vial.

CONTINUE WITH
HANDLING CHECK (NEXT PAGE).

Please insert a new test strip and set the
code as indicated on your test strip vial.



PROBLEM: ASSUMED ABNORMAL VALUE

A: HANDLING CHECK





PROBLEM: ASSUMED ABNORMAL VALUE

B: DEVICE CHECK




C: TEST STRIP CHECK




PROBLEM: ASSUMED ABNORMAL VALUE

D: DEVICE AND TEST STRIP CHECK



PROBLEM: DIFFERENT VALUE THAN REFERENCE DEVICE

E: DIFFERENT VALUE THAN COMPETITOR HANDHELD METER




DO NOT COMPARE TWO HANDHELD METERS!

We cannot determine whether one or both devices are wrong or both devices show the right result within the tolerance range. The correct result might also be in between.

IF YOU WANT TO COMPARE THE RESULTS WITH ANOTHER METER, PLEASE USE A LABORATORY DEVICE IN THE CORRECT WAY ACCORDING TO SECTION F: LABORATORY DEVICE

TO CHECK THE CORRECT FUNCTIONALITY OF YOUR DEVICE, PLEASE SEE ALSO SECTION B: DEVICE

F: DIFFERENT VALUE THAN LABORATORY DEVICE



Laboratory device (name, manufacturer):

Were both samples taken at the same time and from the same puncture site?

NO → Repeat the measurement with a blood sample taken at the same time with the same puncture site.

YES ↓

Were both samples taken from the same blood sample (both capillary blood)?

NO → Repeat the measurement with the same blood sample.

YES ↓

What type of blood was used for the lab device?

Plasma.

Plasma is okay, since Omnitest® plus is calibrated against plasma.

Whole Blood.

Wrong. Repeat the reference measurement with plasma blood sample as Omnitest® plus is plasma-calibrated and can only be compared with plasma blood sample of a lab device.

Was the blood sample centrifuged immediately after collecting the blood sample (within 10 min.)?

NO → Was the natural glycolysis inhibited? → NO → Repeat the measurement with a new blood sample. The glycolysis may lead to a wrong low reading.

YES ↓

What are the measured values of both devices?
Laboratory device: _____ Omnitest® plus: _____

We assume that the laboratory device test result is correct. Please compare: Is the Omnitest® plus result within the ISO standard?

Blood glucose level	Tolerance range
≥ 100 mg/dL (5.55 mmol/L)	± 15%
< 100 mg/dL (5.55 mmol/L)	± 15 mg/dL (0.83 mmol/L)

YES ↓ THE OMNITEST® PLUS DEVICE IS WORKING PROPERLY

NO ↓ THE OMNITEST® PLUS DEVICE MAY BE INACCURATE. PLEASE FOLLOW THE INSTRUCTIONS IN SECTION A: HANDLING